

Caution in Usage of Synthetic Mesh in Repair of Incisional Hernia

Ahmad Izadpanah^{1,*}

¹ Colorectal Research Center, Shiraz University of Medical Sciences, Shiraz, IR Iran

*Corresponding author: Ahmad Izadpanah, Colorectal Research Center, Shiraz University of Medical Sciences, Shiraz, IR Iran. Tel.: +98-7112306972, Fax:+98-7112330724, E-mail: iza-dpana@sums.ac.ir

Received: Jun 29, 2013; Revised: Jun 30, 2013; Accepted: July 05, 2013

Keywords: Ventral incisional hernia; Hernia repair; Mesh graft

Dear Editor,

Variety of mesh materials are available, for treatment of ventral hernia repair, that can be selected by surgeons. Three main types of prosthetic mesh are used for repair of abdominal hernias (1-4).

These types are:

1- Synthetic mesh, such as polypropylene (pp) or poly ester with vigorous tissue ingrowth and high tensile strength.

2- Biologic mesh is a collagen based human, bovine or porcine scaffold (4, 5). This type of mesh is used in the setting of contaminated or infected surgical incisions; and can be implanted intra or extra-peritoneal position.

3- Composite, or barrier-coated is a dual sided prosthetic having a visceral surface that repels tissue in growth and parietal side to provide a strong repair(4). The visceral side decreases adhesive formation. Synthetic mesh such as polyester are appropriate durable materials for extra peritoneal placement(4-6). And if we used intraperitoneally could develop severe adhesions to the bowel and causes enterocutaneous fistula (2, 7). This phenomena is confirmed by many studies. Even mesh related fistula or their complications have seen in these patients who were treated using composite dual sided prosthetic meshes (2, 3).

In the study done by Hosseini et al. (8) nylon type mesh is used, which is in 1st group and should not be used intra peritoneally. How the author did not see any complication is questionable (8). In underlay method it is better to put the mesh on the peritoneal layer, beneath the transversalis fascia to prevent mesh complications such as bowel fistula (3-5). It may be due to short follow up time. But

their using method could be useful in patients with ventral hernia and patients with thin or distracted fascia. Because can anchor the appropriating stitches in two sided implanted mesh.

Acknowledgements

There is no acknowledgment.

Financial Disclosure

There is no financial disclosure.

References

1. Zhang Z, Zhang T, Li J, Ji Z, Zhou H, Zhou X, et al. Preparation of poly(L-lactic acid)-modified polypropylene mesh and its antiadhesion in experimental abdominal wall defect repair. *J Biomed Mater Res B Appl Biomater*. 2013.
2. Sommer T, Friis-Andersen H. DynaMesh in the repair of laparoscopic ventral hernia: a prospective trial. *Hernia*. 2013.
3. Wiegner A, Schlegel N, Isbert C, Jurowich C, Dohr S, Germer CT, et al. Lessons and challenges during a 5-year follow-up of 21 Compositix Kugel implantations. *Hernia*. 2013.
4. Cevasco M, Itani KM. Ventral hernia repair with synthetic, composite, and biologic mesh: characteristics, indications, and infection profile. *Surg Infect (Larchmt)*. 2012;13(4):209-15.
5. Kissane NA, Itani KM. A decade of ventral incisional hernia repairs with biologic acellular dermal matrix: what have we learned? *Plast Reconstr Surg*. 2012;130(5 Suppl 2):194s-202s.
6. Mismar A, Al-Ardah M, Albsoul N, Younes N. Underlay mesh repair for spontaneous lumbar hernia. *Int J Surg Case Rep*. 2013;4(6):534-6.
7. Sadava EE, Krpata DM, Gao Y, Rosen MJ, Novitsky YW. Wound healing process and mediators: Implications for modulations for hernia repair and mesh integration. *J Biomed Mater Res A*. 2013.
8. Hosseini Hosseini S, Hosseini SV, Ghahramani L, Rezaianzadeh A, Safarpour AR, Rahimikazerooni S. Modified Sandwich Technique Mesh Implantation in Repair of Incisional Hernia. *Ann Colorectal Res*. 2013;1(1):28-31.